

**NEW
2018**



DIGITAL AUTO CLEVELAND FLASH AND BURNING POINT APPARATUS

Standard: EN 22592, ASTM D92, AASHTO T48, IP 36/67, UNE 7075, NFT60-118, ISO 2592

Determination of Flash & Fire Point

The apparatus is applicable for determining measurement of flash points and burning points of petroleum products (the Cleveland open cup method) except petroleum products with flash points and burning points lower than 79°C.

Technical Specifications :

Model Number	NL 2039 X / 003
Ignition	Automatic
Heating Power of Furnace	≤ 650 W
Temperature Adjustment	Controlled Silicon (Continuous)
Dimension (mm)	265 (L) x 340 (W) x 330 (H)
Approx. Weight	6 kg
Power	220 V, 1 ph, 50 / 60 Hz

Unit Consists Of :

Model Number	Parts Description	Qty
NL 2039 X / 003 – P 001	Thermocouple Sensor (K Type)	1no.
NL 2039 X / 003 – P 002	Brass Cup With Handle	1no.
NL 2039 X / 003 – P 003	Flame Gas Device Unit c/w Gas Hose	1 unit.

*1 Copy of Manual Instruction

Optional Accessories :

Model Number	Accessories Description
NL 2039 X / 002 – A 001	Flame Gas Regulator