



**NEW
2018**

MORTAR WORKABILITY APPARATUS

Standard: EN 413-2/459-2, NF P18-452

Workability Of Masonry Cement & Building Limes Test

Used to verify the homogeneity of concrete in relation to its workability or plasticity. This test method has particular applications for concretes containing chemical admixtures. The apparatus consists of a metal box divided into two parts and fitted with an electrical vibrator system.

Technical Specifications :

Model Number	NL 3008 X / 001
Dimension (mm)	230 (L) x 155 (W) x 240 (H)
Controller Dimension (mm)	150 (L) x 125 (W) x 340 (H)
Approx. Weight	18.5 kg
Power	220 - 240 V, 50 / 60 Hz, 1 Phase

Unit Consists Of :

Model Number	Parts Description	Qty
NL 3008 X / 001 - P001	Mortar Compression Rod	1no.

Optional Accessories :

Model Number	Parts Description
NL 7031 X / 001	Digital Stop Watch